



Certificate of Compliance

Certificate: 1759607

Master Contract: 205557

Project: 80012395

Date Issued: 2019-09-16

Issued To: Endress+Hauser Conducta GmbH & Co. KG
Dieselstraße 24
Gerlingen, Baden-Württemberg, 70839
Germany

Attention: Robert Binder

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: *Dimcho Genov*
Dimcho Genov



PRODUCTS

CLASS - C225206 - PROCESS CONTROL EQUIPMENT

CLASS - C225286 - PROCESS CONTROL EQUIPMENT Certified to US Standards

Two-wire transmitter Liquiline M for digital sensors:

Model CM42-XXXXXXXXXX, rated max 30V dc and 32V dc for PA/FF, max 5.3 W, maximum ambient 70 °C;
Enclosure type 4X; (X=A-Z or 0-9 denote non-safety relevant indications)



Certificate: 1759607
Project: 80012395

Master Contract: 205557
Date Issued: 2019-09-16

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12

ANSI/UL Std. No. 61010-1 (3rd Edition)

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements
- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 1759607

Master Contract: 205557

***The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.***

Product Certification History

Project	Date	Description
80012395	2019-09-16	Update report 1759607 to cover revisions in the enclosure construction
70217338	2019-04-23	Update report 1759607 to cover the removal of model CM330
000070012534	2014-10-22	Supersedes report 1759607. Update report to cover evaluation to the 3rd edition of standard 61010-1, revising of CM42 ratings and making other minor report changes.